

AUXIER & ASSOCIATES, INC.

PAP-KAN

1428

**STANDARD LEVEL IV
REPORT OF ANALYSIS**

WORK ORDER #15-10128-OR

October 30, 2015

**EBERLINE ANALYTICAL/OAK RIDGE LABORATORY
OAK RIDGE, TN**

TABLE OF CONTENTS

SECTION	DESCRIPTION	PAGE
I	Chain of Custody & pH Check Sheet	0004
II	Sample Acknowledgement	0008
III	Case Narrative	0011
IV	Analytical Results Summary	0014
V	Analytical Standards	0016
VI	Quality Control Sample Results Summary	0021
VII	Laboratory Technician's Notes & Runlogs	0024
VIII	Analytical Data (Gross Alpha/Beta)	0029
	Last Page Number	0048



**Eberline Services – Oak Ridge Laboratory
LABORATORY DATA SUPPORT CHECKLIST**


MP-001-3

Eberline Services Work Order # 15 - 10 128

The checklist items listed below are to be initialed by appropriate staff upon completion/verification.

Date for Partial	Initials	Date	Initials	Checklist Items
		10/21/15	JED	Sample Log-In
		10/23/15	EJT	Data Compilation
		10/29/15	MLT	First Technical Data Review
		10/29/15	USA	Second Technical Data Review
		10/28/15	EJT	Data Entry/Electronic Deliverable
		10/28/15	EJT	Case Narrative
		10/30/15	RBS	Electronic Deliverable Proof
		10/30/15	USA	Samples Analyzed within Holding Time Yes? <input checked="" type="checkbox"/> No? <input type="checkbox"/>
		10/30/15	USA	QA/QC Review
		10/23/15	EJT	Client in Possession of Data Electronic or Hard Copy
				Invoiced by Laboratory

Technical/Clerical Corrections, Signatures Needed, Problems, Etc	Date/Initials

Date package approved by:  10/30/15
 Laboratory Manager Date

SECTION I
CHAIN OF CUSTODY
&
pH CHECK SHEET



EBERLINE
SERVICES
Oak Ridge Laboratory

Internal Chain of Custody

Work Order #	15-10128
Lab Deadline	10/26/2015
Analysis	GaGdT_ThSr - Level 4
Sample Matrix	Water

Comments	Sample Fraction	HP 210 / 270 Detector Activity	Storage Location
Re-log of 15-10065 all fractions	04	37	NN1.1
	05	36	NN1.1
	06	36	NN1.1
	07	35	NN1.1
	08	32	NN1.1
	09	36	NN1.1

	Location (circle one)					Initials	Date
Received by	Sample Storage	Rough Prep	Prep	Separations	Count Room	MA	22 OCT 15
Relinquished by	Sample Storage	Rough Prep	Prep	Separations	Count Room	MA	22 OCT 15 0845
Received by	Sample Storage	Rough Prep	Prep	Separations	Count Room		C 10/22/15
Relinquished by	Sample Storage	Rough Prep	Prep	Separations	Count Room	WB	10/22/15 1516
Received by	Sample Storage	Rough Prep	Prep	Separations	Count Room		
Relinquished by	Sample Storage	Rough Prep	Prep	Separations	Count Room		
Received by	Sample Storage	Rough Prep	Prep	Separations	Count Room		
Relinquished by	Sample Storage	Rough Prep	Prep	Separations	Count Room		
Received by	Sample Storage	Rough Prep	Prep	Separations	Count Room		
Relinquished by	Sample Storage	Rough Prep	Prep	Separations	Count Room		
Received by	Sample Storage	Rough Prep	Prep	Separations	Count Room		
Relinquished by	Sample Storage	Rough Prep	Prep	Separations	Count Room		

	Sample Receiving Report (Volumes, pH, & CPM)	Internal Work Order
		15-10128
		Received By JBAILEY

FR	ClientID	# Btls	Comments	Matrix	Storage	Rec Vol Ttl	CPM Max	
01	LCS	0		WA	NN1.1			
02	BLANK	0		WA	NN1.1			
03	DUP	0		WA	NN1.1			
04	KC91-159-L ✓	1		WA	NN1.1	3.76	37	
				Container Number:	pH Orig	pH Final	Volume (L)	CPM
				1	7	7	3.7600	37
05	KC91-159-U ✓	1		WA	NN1.1	3.76	36	
				Container Number:	pH Orig	pH Final	Volume (L)	CPM
				1	7	7	3.7600	36
06	OW-008-L ✓	1		WA	NN1.1	3.76	36	
				Container Number:	pH Orig	pH Final	Volume (L)	CPM
				1	7	7	3.7600	36
07	KC-90-137-U ✓	1		WA	NN1.1	3.76	35	
				Container Number:	pH Orig	pH Final	Volume (L)	CPM
				1	7	7	3.7600	35
08	KC90-137-L ✓	1		WA	NN1.1	3.76	32	
				Container Number:	pH Orig	pH Final	Volume (L)	CPM
				1	7	7	3.7600	32
09	RB-01-09 21 15 ✓	1		WA	NN1.1	0.75	36	
				Container Number:	pH Orig	pH Final	Volume (L)	CPM
				1	7	7	0.7500	36

*Key
10/21/15*

Received by: *James Bailey* Date: 10-21-15

SECTION II
SAMPLE ACKNOWLEDGEMENT



STANDARD OPERATING PROCEDURE

Sample Receiving

MP-001, Rev. 15
Effective: 2/2/15
Page 13 of 15

Eberline Services – Oak Ridge Laboratory

SAMPLE RECEIPT CHECKLIST MP-001-2

WORK ORDER # 15 - 10 128

SAMPLE MATRIX/MATRICES:

(CIRCLE ONE OR BOTH)

AQUEOUS NON-AQUEOUS

(CIRCLE EITHER YES, NO, OR N/A)

WERE SAMPLES:

Received in good condition?	<input checked="" type="radio"/> Y	N	
If aqueous, properly preserved	<input checked="" type="radio"/> Y	N	N/A

WERE CHAIN OF CUSTODY SEALS:

Present on outside of package?	<input checked="" type="radio"/> Y	N
Unbroken on outside of package?	<input checked="" type="radio"/> Y	N
Present on samples?	<input checked="" type="radio"/> Y	N
Unbroken on samples?	<input checked="" type="radio"/> Y	N
Was chain of custody present upon sample receipt?	<input checked="" type="radio"/> Y	N

IF THE RESPONSE TO ANY OF THE ABOVE IS NO, A DISCREPANT SAMPLE RECEIPT REPORT (DSR) HAS BEEN ISSUED.

REMARKS: _____

SIGNATURE:  DATE: 10-21-15

SECTION III
CASE NARRATIVE



EBS-OR-39887

October 30, 2015

Cecilia Greene
USA ENV LP/Auxier & Associates, Inc.
9821 Cogdill Road, Suite 1
Knoxville, TN 37932

CASE NARRATIVE
Work Order# 15-10128-OR

SAMPLE RECEIPT

This work order contains six water samples received 10/12/2015 and re-logged at the client's request 10/21/2015. These samples were analyzed for Gross Alpha/Beta.

<u>CLIENT ID</u>	<u>LAB ID</u>	<u>CLIENT ID</u>	<u>LAB ID</u>
KC91-159-L	15-10128-04	KC-90-137-U	15-10128-07
KC91-159-U	15-10128-05	KC90-137-L	15-10128-08
OW-008-L	15-10128-06	RB-01-09 21 15	15-10128-09

ANALYTICAL METHODS

Gross Alpha/Beta was analyzed using EPA Method 900.0 Modified.

ANALYTICAL RESULTS

Combined Standard Uncertainty is reported at 2-sigma value.

Minimum Detectable Activity (MDA) values for data represented in this report are sample-specific. MDA measurements are determined based on factors and conditions including instrument settings, aliquot size and matrix type.

GROSS ALPHA & BETA

Samples were prepared by evaporation of representative volumetric aliquots acidified with HNO₃. Reduced samples were then transferred to steel planchets for final evaporation to dryness and flaming. Samples were then counted on a gas proportional counter. Results were corrected as required for inherent self-absorption based on residual mass present.


ANALYTICAL RESULTS CONTINUED

GROSS ALPHA & BETA CONTINUED

Samples demonstrated acceptable results for all Gross Alpha and Beta analyses. The Gross Alpha and Beta method blank demonstrated results slightly greater than the detection limit. In each case blank results were background equivalent. Results for the Gross Alpha and Beta duplicate demonstrated a high relative percent difference; however, normalized difference is within acceptable limits for the analytical technique. Results for the Gross Alpha and Beta laboratory control sample demonstrated an acceptable percent recovery.

CERTIFICATION OF ACCURACY

I certify that this data report is in compliance with the terms and conditions of the Purchase Order, both technically and for completeness, for other than the conditions detailed above. Release of the data contained in this hard copy data package has been authorized by the cognizant project manager or his/her designee to be accurate as verified by the following signature.



M.R. McDougall
Laboratory Manager

Date: 10/30/2015

Eberline Analytical wants and encourages your feedback regarding our performance providing radioanalytical services. Please visit <http://www.eberlineservices.com/client.htm> to provide us with feedback on our services.

SECTION IV
ANALYTICAL RESULTS SUMMARY

Eberline Analytical

Final Report of Analysis

Lab ID		Sample Type	Client ID	Sample Date	Receipt Date	Analysis Date	Batch ID	Analyte	Method	Result	CU	CSU	MDA	CV	Report Units
<p style="text-align: center;">Cecilia Greene Auxier & Associates, Inc. 9821 Cogdill Road, Suite 1 Knoxville, TN 37830</p>															
<p style="text-align: center;">Work Order Details: 15-10128</p>															
<p style="text-align: center;">SDG: PAP-KAN</p>															
<p style="text-align: center;">Purchase Order: ENVIRONMENTAL</p>															
<p style="text-align: center;">Analysis Category: WA</p>															
<p style="text-align: center;">Sample Matrix:</p>															
15-10128-01	LCS	KNOWN		10/21/15 09:00	10/21/2015	10/22/2015	15-10128	Gross Alpha	EPA 900.0 Modified	2.68E+02	1.15E+01				pCi/l
15-10128-01	LCS	SPIKE		10/21/15 00:00	10/21/2015	10/22/2015	15-10128	Gross Alpha	EPA 900.0 Modified	2.80E+02	3.74E+00	3.09E+01	1.87E-01	1.92E-01	pCi/l
15-10128-02	MBL	BLANK		10/21/15 00:00	10/21/2015	10/22/2015	15-10128	Gross Alpha	EPA 900.0 Modified	5.40E-01	2.06E-01	2.13E-01	2.84E-01	2.81E-01	pCi/l
15-10128-03	DUP	KC91-159-L		09/26/15 09:25	10/21/2015	10/22/2015	15-10128	Gross Alpha	EPA 900.0 Modified	8.37E+00	3.09E+00	3.17E+00	4.78E+00	1.54E+00	pCi/l
15-10128-04	DO	KC91-159-L		09/26/15 09:25	10/21/2015	10/22/2015	15-10128	Gross Alpha	EPA 900.0 Modified	8.38E+00	3.47E+00	3.58E+00	5.15E+00	1.37E+00	pCi/l
15-10128-05	TRG	KC91-159-U		09/26/15 10:10	10/21/2015	10/22/2015	15-10128	Gross Alpha	EPA 900.0 Modified	8.38E+00	3.47E+00	3.58E+00	5.15E+00	1.37E+00	pCi/l
15-10128-06	TRG	OW-008-L		09/26/15 10:51	10/21/2015	10/22/2015	15-10128	Gross Alpha	EPA 900.0 Modified	5.37E+00	2.19E+00	2.26E+00	3.02E+00	5.66E-01	pCi/l
15-10128-07	TRG	KC-90-137-U		09/26/15 12:05	10/21/2015	10/22/2015	15-10128	Gross Alpha	EPA 900.0 Modified	9.87E+00	2.57E+00	2.78E+00	2.50E+00	4.56E-01	pCi/l
15-10128-08	TRG	KC90-137-L		09/26/15 12:40	10/21/2015	10/22/2015	15-10128	Gross Alpha	EPA 900.0 Modified	3.72E+00	1.76E+00	1.81E+00	2.70E+00	6.31E-01	pCi/l
15-10128-09	TRG	RB-01-09 21 15		09/21/15 11:20	10/21/2015	10/22/2015	15-10128	Gross Alpha	EPA 900.0 Modified	1.33E-01	5.52E-01	5.53E-01	1.27E+00	1.28E+00	pCi/l
15-10128-01	LCS	KNOWN		10/21/15 00:00	10/21/2015	10/22/2015	15-10128	Gross Beta	EPA 900.0 Modified	2.92E+02	8.75E+00				pCi/l
15-10128-01	LCS	SPIKE		10/21/15 00:00	10/21/2015	10/22/2015	15-10128	Gross Beta	EPA 900.0 Modified	2.77E+02	3.13E+00	3.84E+01	6.50E-01	1.62E+00	pCi/l
15-10128-02	MBL	BLANK		10/21/15 00:00	10/21/2015	10/22/2015	15-10128	Gross Beta	EPA 900.0 Modified	9.45E-01	2.70E-01	3.00E-01	4.45E-01	7.94E-01	pCi/l
15-10128-03	DUP	KC91-159-L		09/26/15 09:25	10/21/2015	10/22/2015	15-10128	Gross Beta	EPA 900.0 Modified	6.79E+00	3.03E+00	3.17E+00	5.57E+00	9.38E+00	pCi/l
15-10128-04	DO	KC91-159-L		09/26/15 09:25	10/21/2015	10/22/2015	15-10128	Gross Beta	EPA 900.0 Modified	1.21E+01	3.76E+00	4.12E+00	6.72E+00	1.34E+01	pCi/l
15-10128-05	TRG	KC91-159-U		09/26/15 10:10	10/21/2015	10/22/2015	15-10128	Gross Beta	EPA 900.0 Modified	1.12E+01	4.10E+00	4.39E+00	7.59E+00	1.69E+01	pCi/l
15-10128-06	TRG	OW-008-L		09/26/15 10:51	10/21/2015	10/22/2015	15-10128	Gross Beta	EPA 900.0 Modified	6.07E+00	1.91E+00	2.08E+00	3.39E+00	5.94E+00	pCi/l
15-10128-07	TRG	KC-90-137-U		09/26/15 12:05	10/21/2015	10/22/2015	15-10128	Gross Beta	EPA 900.0 Modified	8.06E+00	1.84E+00	2.15E+00	2.96E+00	4.60E+00	pCi/l
15-10128-08	TRG	KC90-137-L		09/26/15 12:40	10/21/2015	10/22/2015	15-10128	Gross Beta	EPA 900.0 Modified	5.15E+00	1.51E+00	1.67E+00	2.59E+00	4.12E+00	pCi/l
15-10128-09	TRG	RB-01-09 21 15		09/21/15 11:20	10/21/2015	10/22/2015	15-10128	Gross Beta	EPA 900.0 Modified	6.42E-01	1.12E+00	1.13E+00	2.33E+00	4.66E+00	pCi/l

CU=Counting Uncertainty; CSU=Combined Standard Uncertainty (2-sigma); MDA=Minimal Detected Activity; LCS=Laboratory Control Sample; MBL=Blank; DUP=Duplicate; TRG=Normal Sample; DO=Duplicate Original; CV=Critical Value



EBERLINE
SERVICES

EBERLINE ANALYTICAL CORPORATION
601 SCARBORO ROAD OAK RIDGE, TN 37830 865/481-0683 FAX 865/483-4621

SECTION V
ANALYTICAL STANDARDS

ANALYTICS

QA/QC REVIEWED
Date 4/30/96 Initials WT

Am-4

1380 Seaboard Industrial Blvd.
Atlanta, Georgia 30318 · U.S.A.

Phone (404) 352-8677
Fax (404) 352-2837

CERTIFICATE OF CALIBRATION

Standard Radionuclide Source

52094-416

Am-241 10 mL Liquid in Flame Sealed Vial

This standard radionuclide source was prepared gravimetrically from a calibrated master liquid radionuclide solution source. The master source was calibrated by liquid scintillation counting.

ANALYTICS maintains traceability to the National Institute of Standards and Technology through Measurements Assurance Programs as described in USNRC Reg. Guide 4.15, Revision 1.

ISOTOPE:	Am-241
ACTIVITY (dps):	1.975 E+05
HALF-LIFE:	432.2 years
CALIBRATION DATE:	March 19, 1996 12:00 EST
TOTAL ERROR:	3.0%
SYSTEMATIC ERROR:	2.37%
RANDOM ERROR:	0.63%

10.01177 grams of solution 1M HCl.

P O NUMBER OR3830, Item 1

SOURCE PREPARED BY: Kare O'Brien Beverly
K. O. Beverly, Radiochemist

Q A APPROVED: DM. [Signature] 4-26-96



QUALITY CONTROL PROGRAM
MP-009

Rev.8; 1/10/03
Title: Radioactive Reference Standards Solutions & Records

EBERLINE SERVICES - OAK RIDGE LABORATORY
RADIOACTIVE REFERENCE STANDARD SOLUTIONS
SECONDARY DILUTION (RE-CERTIFICATION)

Solution Reference # Analytix 52094-416 Date 8/5/2015 0:00
Solution # A/B-7 (alpha)

Principal Radionuclide ²⁴¹Am Half Life, Years 4.322E+02 Half Life, Days 1.579E+05

Radionuclide of Interest ²⁴¹Am Reference Date 3/19/1996 0:00
Parent Solution Conc. 1.19E+04 dpm/ml

Chemical Composition of Standard Solution
²⁴¹AmCl₃ in 1M HCL

Dilution Instructions: Dilution Solvent Used 1M HNO₃

SECONDARY VOLUMETRIC DILUTION

Vol. Parent Solution: 60.0000 ml
Total Activity: 7.1100E+05 dpm Final Activity Concentration: 7.1100E+02 dpm/ml
Final Volume: 1000.00 ml

NOTES:

This activity concentration is based on the original reference date listed above. All activities are corrected to the date and time of analysis by the laboratory data processing software.

Expiration Date: August 4, 2016

Verified & Approved By: [Signature]

Date: 8/5/15

QC Approval: [Signature]

Date: 8/5/15

SR-42
13 uhh



National Institute of Standards & Technology Certificate

Standard Reference Material 4234A Strontium-90 Radioactivity Standard

This Standard Reference Material (SRM) consists of radioactive strontium-90 chloride, non-radioactive strontium chloride, non-radioactive yttrium chloride, and hydrochloric acid dissolved in 5 mL of distilled water. The solution is contained in a flame-sealed NIST borosilicate-glass ampoule. The SRM is intended for the calibration of beta-particle counting instruments and for the monitoring of radiochemical procedures.

Radiological Hazard

The SRM ampoule contains strontium-90 with a total activity of approximately 13 MBq. Strontium-90 decays by beta-particle emission to yttrium-90, which also decays by beta-particle emission. None of the beta particles escape from the SRM ampoule. The beta particles emitted from strontium-90 and yttrium-90 produce bremsstrahlung photons with energies up to 2 MeV. Most of these photons escape from the SRM ampoule and can represent a radiation hazard. Approximate unshielded dose rates at several distances (as of the reference time) are given in note [a]*. Appropriate shielding and/or distance should be used to minimize personnel exposure. The SRM should be used only by persons qualified to handle radioactive material.

Chemical Hazard

The SRM ampoule contains hydrochloric acid (HCl) with a concentration of 1 mole per liter of water. The solution is corrosive and represents a health hazard if it comes in contact with eyes or skin. If the ampoule is to be opened to transfer the solution, the recommended procedure is given on page 2. The ampoule should be opened only by persons qualified to handle both radioactive material and strong acid solution.

Storage and Handling

The SRM should be stored and used at a temperature between 5 and 65 °C. The solution in an unopened ampoule should remain stable and homogeneous until at least March 2005.

The ampoule (or any subsequent container) should always be clearly marked as containing radioactive material. If the ampoule is transported it should be packed, marked, labeled, and shipped in accordance with the applicable national, international, and carrier regulations. The solution in the ampoule is a dangerous good (hazardous material) both because of the radioactivity and because of the strong acid.

Preparation

This Standard Reference Material was prepared in the Physics Laboratory, Ionizing Radiation Division, Radioactivity Group, J.M.R. Hutchinson, Group Leader. The overall technical direction and physical measurements leading to certification were provided by L.L. Lucas of the Radioactivity Group and D.B. Golas, Nuclear Energy Institute Research Associate.

The support aspects involved in the preparation, certification, and issuance of this SRM were coordinated through the Standard Reference Materials Program by N.M. Trahey.

Gaithersburg, Maryland 20899
May 1995 (Text only revised November 1997)

Thomas E. Gills, Chief
Standard Reference Materials Program



QUALITY CONTROL PROGRAM
QCP-009

Rev.7: 9/29/99
Title: Radioactive Reference Standards Solutions & Records

EBERLINE SERVICES - OAK RIDGE LABORATORY
RADIOACTIVE REFERENCE STANDARD SOLUTIONS
SECONDARY DILUTION (RE-CERTIFICATION)

Solution Reference #		QCP-009-1-A	Date	8/5/2015 0:00
Principal Radionuclide		NIST 4234A	Solution #	A/B-7 (beta)
	Half Life, Years	2.878E+01	Half Life, Days	1.051E+04
	⁹⁰ Strontium			

Radionuclide of Interest	⁹⁰ Sr	Reference Date	3/13/1995 0:00
Parent Solution Conc.	1.52E+06 dpm/ml		

The beta activity of solution reflects the original ⁹⁰Strontium concentration and an equal concentration of ⁹⁰Yttrium.

Chemical Composition of Standard Solution
⁹⁰SrCl₂ in 1 M HCl

Dilution Instructions:	Dilution Solvent Used	1 M HNO ₃
------------------------	-----------------------	----------------------


SECONDARY VOLUMETRIC DILUTION


Vol. Parent Solution:	0.5000 ml	Final Activity Concentration:	7.5764E+02 dpm/ml
Total Activity:	7.5764E+05 dpm		
Final Volume:	1000.00 ml		

This activity concentration is based on the original reference date listed above. All activities are corrected to the date and time of analysis by the laboratory data processing software.

NOTES:

Expiration Date: August 4, 2016

Verified & Approved By:  Date: 08/05/15

QC Approval:  Date: 8/5/15

SECTION VI
QUALITY CONTROL SAMPLE RESULTS SUMMARY

W/O	Analysis	Run	Activity Units	Aliquot Units	Client Name
15-10128	GaGbT_ThSr	1	pCi	I	Auxier & Associates, Inc.

Laboratory Control Sample

Analyte	LCS Measured	CSU Measured	LCS Expected	Uncert. Expected	Known MS CSU	Known Error	Result	CSU	Standard ID	Standard ACT (dpm)	Standard Error	Standard Added (g)
GROSS ALPHA_TH	104.81%	11.00%	100.00%	4.30%	2.68E+02	1.15E+01	2.80E+02	3.09E+01	A/B-07	5.96E+02	4.30E+00	9.97E-01
GROSS BETA_SR	95.04%	13.86%	100.00%	3.00%	2.92E+02	8.75E+00	2.77E+02	3.84E+01	A/B-07	6.49E+02	3.00E+00	9.97E-01

Matrix Spike

Analyte	Normalized Difference	MS Actual % Rec	Expected MS Result	Expected MS Uncert	Actual MS Result	Actual MS CSU	Sample Result	Sample CSU	Sample Aliquot	Standard ID	Standard ACT (dpm)	Standard Error %	Standard Added (g)

Replicate Sample


Analyte	Normalized Difference	RPD	Original Result	Original CSU	Replicate Result	Replicate CSU	LCS Relative Bias	LCS % R	MS % R	MS ND	Rep RPD	Rep ND
GROSS ALPHA_TH	1.07	33.40	8.93E+00	3.47E+00	6.37E+00	3.17E+00	1.05	OK			NA	OK
GROSS BETA_SR	1.99	56.04	1.21E+01	4.12E+00	6.79E+00	3.17E+00	0.95	OK			NA	OK

QC Summary


Analyte	Normalized Difference	RPD	Original Result	Original CSU	Replicate Result	Replicate CSU	LCS Relative Bias	LCS % R	MS % R	MS ND	Rep RPD	Rep ND
GROSS ALPHA_TH	1.07	33.40	8.93E+00	3.47E+00	6.37E+00	3.17E+00	1.05	OK			NA	OK
GROSS BETA_SR	1.99	56.04	1.21E+01	4.12E+00	6.79E+00	3.17E+00	0.95	OK			NA	OK

WO	Analysis	Run	Activity Units	Aliquot Units	Client Name
15-10128	GaGbT_ThSr	1	pCi	I	Auxier & Associates, Inc.


LCS % Recovery



Replicate Sample RPD




Normalized Difference




No Matrix Spike

SECTION VII
LABORATORY TECHNICIAN'S NOTES
& RUNLOGS

 EBERLINE <small>SERVICES</small> Work Order Analysis Notes	Oak Ridge Laboratory 601 Scarboro Rd. Oak Ridge, TN 37830 Voice: 865.481.0683 www.eberlineservices.com		Internal Work Order	15-10128
			Analysis Code	GaGbT_ThSr
			Run Number	1

#	Date	Dept	User	Notes
1	10/22/15 03:40	PREP	MHIGHTOWER	Aliquots based on previous TDS (15-10065).
2	10/22/15 03:40	PREP	MHIGHTOWER	Aliquoted samples, dried, nitrated, transferred to tared planchets, dried, flamed, re-weighed, and submitted to count room

Mh 22 OCT 15

 <p>EBERLINE SERVICES</p> <p>Reagents Used in an Analysis</p>		Internal Work Order			
		15-10128			
		Analysis Code		Run	
		GaGbT_ThSr		1	
Reagent ID	Reagent Name	Reagent Concentration	Analyst ID	Date Recorded	
016403D12	Nitric Acid	3N	MHIGHTOWER	10/22/2015	

Date	Sample	Client	Time	Other	Analysis	Test
10/15	GFPRC	LAB	0506	✓	LAB	✓
10/15	Buwalda	LAB	0542	✓	LAB	✓
10/15	151000754(1-4-6)	UCON	0708	✓	SPEOLY	✓
10/15	1510081AD(2-3)	Parsons	0717	Blr	LAB	✓
10/15	1510054(14-2-6)	USt	0978	✓ HSE	LAB	✓
10/15	Daily Bkgd	Lab	0609	140p	OR	AC
10/15	ZFAL1020	LAB	0624	✓	LAB	✓
10/15	1510071NP(1-4)	LAB	0710	✓	LAB	✓
10/15	1510071NP(1-4)	UCON	0847	✓	NP272	✓
10/15	1510021NP(1-7)	UCON	0852	✓	NP272	✓
10/15	150809754(2-4)	POB	0921	✓	SPEOLY	✓
10/15	1510030(14-10-1)	STOPND	0959	✓	LAB	✓
10/15	1510030(14-10-1)	STOPND	0919	✓	LAB	✓
10/15	1508151(14-12-4)	UCON	1010	✓	LAB	✓
10/120	150909754(1-1)	DOB	1101	✓	SPEOLY	✓
10/120	150909754(1-29)	DOB	1239	✓	SPEOLY	✓
10/20/16	RA SD rect - (1-5)	Lab	1219	15min	Ra	KB
10/21	GFPRC	LAB	0518	✓	LAB	✓
10/21	Buwalda	LAB	0554	✓	LAB	✓
10/21	150805654(1-6)	POB	0737	✓	SPEOLY	✓
10/21	1508150M(1-7)	Account	0944	✓	LAB	✓
10/21	1510001P(1-7)	Slaway	1714	✓	P2710	✓
10/22	GFPRC	LAB	0515	✓	LAB	✓
10/22	Buwalda	LAB	0550	✓	LAB	✓
10/22	1510051MP(1-3-5b)	TM Rpt.	0823	✓	NP272	✓
10/22	1510058MP(1-5)	USteal	0872	✓	NP272	✓
10/22	1510115R(1-5)	UCON	0850	✓	NP207	✓
10/22	1510128AD(2-4)	Auxier	0970	✓	LAB	✓
10/22	1510002M(2-6)	Account	1055	✓	LAB	✓
10/22	1510021R(1-2-4)	UCON	1177	✓	LAB	✓
10/22	1510021R(1-2-4) on (1-7-5)	Auxier		✓	LAB	✓
10/22	1510123AD(1-7-5)	Auxier	1700	✓	LAB	✓
10/22	1510128AD(1-6)	Auxier	1714	✓	LAB	✓

CB4110 Aqua

Rate	Amplified	Client	PodTime	CTD	Insights	Prod
1012	1510058NA(2)	United	0822	14	NP27	✓
1012	151008115(N)	Austria	0922	21	AB	✓

SECTION VIII
ANALYTICAL DATA (GROSS ALPHA/BETA)

Internal Work Order		Run	Analysis Code		Date	Technician		Technician Initials		Witness Initials		
15-10128		1	GaGbT_ThSr		10/22/2015 4:12	MHIGHTOWER		M				
LCS & Matrix Spikes												
Isotope	Sol #	Activity dpm/g	Solution Date	Approx Addition	LCS Volume Used (g)	MS Volume Used (g)	LCS Volume Used (g)	MSD Volume Used (g)	LCS Known pCi	MS Error Estimate	LCS Error Estimate	MSD Error Estimate
Am-241	A/B-07	595.990	10/22/2015	0.790	0.9968		267.60	11.507	0.00	0.000	0.000	0.000
SrY-90	A/B-07	649.300	10/22/2015	0.850	0.9968		291.54	8.746	0.00	0.000	0.000	0.000
1C-99 MS 1C-2a 22043.636 7/5/2014 0.1 Tracers												
fraction	Isotope	Sol #	Activity dpm/g	Solution Date	Volume Used (g)	Approx Addition	Tracer					LCS
							Matrix Spike					

040000

10/22/2015
10/22/2015

Detector ID	Sample ID	Alpha	Beta	Count Time	Voltage	TOD
G2	1510128-01	21567	40871	120	1410	10/22/2015 1:13:58 PM
G4	1510128-06	42	308	120	1410	10/22/2015 1:13:58 PM

(2) 10/22/15
140

Detector ID	Sample ID	Alpha	Beta	Count Time	Voltage	TOD
F1	1510128-01	9	175	120	1410	10/22/2015 9:30:17 AM
F2	1510128-02	50	227	120	1410	10/22/2015 9:30:17 AM
F4	1510128-03	38	212	120	1410	10/22/2015 9:30:17 AM
A2	1510128-04	47	332	120	1410	10/22/2015 9:32:14 AM
A3	1510128-05	49	367	120	1410	10/22/2015 9:32:14 AM
E2	1510128-07	73	327	120	1410	10/22/2015 1:00:28 PM
E3	1510128-08	39	260	120	1410	10/22/2015 1:00:28 PM
E4	1510128-09	10	138	120	1410	10/22/2015 1:00:28 PM

GPC Detector Report
(ALL Backgrounds)

1022

Detector	Alpha/Beta	Calibration Date	Count Date	Bkg CPM	PFW	LCL	Mean	UCL
LB4110A - A1	Alpha	10/20/2015	10/22/2015	2.33E-01	P	-1.11E-01	1.00E-01	3.11E-01
LB4110A - A2	Alpha	10/20/2015	10/22/2015	6.67E-02	P	-3.46E-02	7.33E-02	1.81E-01
LB4110A - A3	Alpha	10/20/2015	10/22/2015	1.00E-01	P	6.62E-02	1.13E-01	1.60E-01
LB4110A - A4	Alpha	10/20/2015	10/22/2015	6.67E-02	P	3.19E-02	1.20E-01	2.08E-01
LB4110A - B1	Alpha	10/20/2015	10/22/2015	8.33E-02	P	1.98E-02	1.23E-01	2.27E-01
LB4110A - B2	Alpha	10/20/2015	10/22/2015	1.17E-01	P	8.41E-02	1.03E-01	1.23E-01
LB4110A - B3	Alpha	10/20/2015	10/22/2015	5.00E-02	P	-1.23E-02	1.17E-01	2.46E-01
LB4110A - B4	Alpha	10/20/2015	10/22/2015	1.17E-01	P	-9.90E-02	9.67E-02	2.92E-01
LB4110A - C1	Alpha	10/20/2015	10/22/2015	2.83E-01	P	-9.17E-02	1.37E-01	3.65E-01
LB4110A - C2	Alpha	10/20/2015	10/22/2015	2.33E-01	P	-4.83E-02	1.80E-01	4.08E-01
LB4110A - C3	Alpha	10/20/2015	10/22/2015	2.83E-01	P	-6.07E-02	1.50E-01	3.61E-01
LB4110A - C4	Alpha	10/20/2015	10/22/2015	2.00E-01	P	-1.20E-01	7.33E-02	2.67E-01
LB4110A - D1	Alpha	10/20/2015	10/22/2015	6.67E-02	P	4.31E-02	1.20E-01	1.97E-01
LB4110A - D2	Alpha	10/20/2015	10/22/2015	1.00E-01	P	4.15E-02	1.07E-01	1.72E-01
LB4110A - D3	Alpha	10/20/2015	10/22/2015	3.33E-02	P	-1.02E-01	8.67E-02	2.75E-01
LB4110A - D4	Alpha	10/20/2015	10/22/2015	1.67E-02	P	-1.25E-02	7.67E-02	1.66E-01
LB4110R - E1	Alpha	10/20/2015	10/22/2015	2.00E-01	P	-1.49E-01	1.25E-01	3.99E-01
LB4110R - E2	Alpha	10/20/2015	10/22/2015	5.00E-02	P	1.13E-02	4.17E-02	7.21E-02
LB4110R - E3	Alpha	10/20/2015	10/22/2015	8.33E-02	P	-2.26E-01	2.00E-01	6.26E-01
LB4110R - E4	Alpha	10/20/2015	10/22/2015	6.67E-02	P	-4.96E-02	4.17E-02	1.33E-01
LB4110R - F1	Alpha	10/20/2015	10/22/2015	6.67E-02	P	6.67E-02	6.67E-02	6.67E-02
LB4110R - F2	Alpha	10/20/2015	10/22/2015	8.33E-02	P	-1.15E-01	1.58E-01	4.32E-01
LB4110R - F3	Alpha	10/20/2015	10/22/2015	2.83E-01	P	2.83E-01	2.83E-01	2.83E-01
LB4110R - F4	Alpha	10/20/2015	10/22/2015	8.33E-02	P	5.86E-03	6.67E-02	1.27E-01
LB4110R - G1	Alpha	10/20/2015	10/22/2015	3.67E-01	F	-9.82E-02	2.67E-01	6.32E-01
LB4110R - G2	Alpha	10/20/2015	10/22/2015	3.33E-02	P	-7.70E-02	7.50E-02	2.27E-01
LB4110R - G3	Alpha	10/20/2015	10/22/2015	3.33E-01	F	2.95E-01	3.25E-01	3.55E-01
LB4110R - G4	Alpha	10/20/2015	10/22/2015	6.67E-02	P	6.67E-02	6.67E-02	6.67E-02

GPC Detector Report
(ALL Backgrounds)

Detector	Alpha/Beta	Calibration Date	Count Date	Bkg CPM	PFW	LCL	Mean	UCL
LB4110A - A1	Beta	10/20/2015	10/22/2015	1.58E+00	P	9.96E-01	1.44E+00	1.88E+00
LB4110A - A2	Beta	10/20/2015	10/22/2015	1.48E+00	P	7.67E-01	1.44E+00	2.11E+00
LB4110A - A3	Beta	10/20/2015	10/22/2015	1.87E+00	P	9.41E-01	1.68E+00	2.42E+00
LB4110A - A4	Beta	10/20/2015	10/22/2015	6.13E+00	F	4.81E+00	5.59E+00	6.37E+00
LB4110A - B1	Beta	10/20/2015	10/22/2015	1.75E+00	P	1.55E+00	1.76E+00	1.98E+00
LB4110A - B2	Beta	10/20/2015	10/22/2015	1.67E+00	P	8.70E-01	1.43E+00	2.00E+00
LB4110A - B3	Beta	10/20/2015	10/22/2015	1.17E+00	P	9.70E-01	1.53E+00	2.10E+00
LB4110A - B4	Beta	10/20/2015	10/22/2015	1.50E+00	P	1.05E+00	1.77E+00	2.48E+00
LB4110A - C1	Beta	10/20/2015	10/22/2015	1.68E+00	P	1.12E+00	1.48E+00	1.83E+00
LB4110A - C2	Beta	10/20/2015	10/22/2015	1.32E+00	P	8.16E-01	1.38E+00	1.95E+00
LB4110A - C3	Beta	10/20/2015	10/22/2015	2.43E+00	F	4.79E-01	1.66E+00	2.84E+00
LB4110A - C4	Beta	10/20/2015	10/22/2015	1.27E+00	P	1.22E+00	1.29E+00	1.36E+00
LB4110A - D1	Beta	10/20/2015	10/22/2015	1.65E+00	P	9.55E-01	1.50E+00	2.04E+00
LB4110A - D2	Beta	10/20/2015	10/22/2015	5.15E+00	F	3.68E+00	4.74E+00	5.80E+00
LB4110A - D3	Beta	10/20/2015	10/22/2015	4.20E+00	F	3.07E+00	5.22E+00	7.37E+00
LB4110A - D4	Beta	10/20/2015	10/22/2015	7.55E+00	F	3.05E+00	9.16E+00	1.53E+01
LB4110R - E1	Beta	10/20/2015	10/22/2015	1.52E+00	P	8.97E-01	1.38E+00	1.87E+00
LB4110R - E2	Beta	10/20/2015	10/22/2015	1.05E+00	P	9.34E-01	1.03E+00	1.12E+00
LB4110R - E3	Beta	10/20/2015	10/22/2015	1.00E+00	P	-3.24E-01	1.50E+00	3.32E+00
LB4110R - E4	Beta	10/20/2015	10/22/2015	1.00E+00	P	5.35E-01	9.00E-01	1.26E+00
LB4110R - F1	Beta	10/20/2015	10/22/2015	1.45E+00	P	8.69E-01	1.33E+00	1.78E+00
LB4110R - F2	Beta	10/20/2015	10/22/2015	7.67E-01	P	-2.27E-01	1.14E+00	2.51E+00
LB4110R - F3	Beta	10/20/2015	10/22/2015	1.92E+00	P	7.16E-01	1.66E+00	2.60E+00
LB4110R - F4	Beta	10/20/2015	10/22/2015	1.02E+00	P	7.08E-01	1.13E+00	1.56E+00
LB4110R - G1	Beta	10/20/2015	10/22/2015	1.72E+00	P	1.41E+00	1.65E+00	1.89E+00
LB4110R - G2	Beta	10/20/2015	10/22/2015	1.62E+00	P	1.37E+00	1.71E+00	2.04E+00
LB4110R - G3	Beta	10/20/2015	10/22/2015	1.58E+00	P	1.21E+00	1.73E+00	2.24E+00
LB4110R - G4	Beta	10/20/2015	10/22/2015	1.35E+00	P	5.36E-01	1.18E+00	1.81E+00

10/22

GPC Detector Report
(ALL Efficiencies)

10/22

Detector	Alpha/Beta	Calibration Date	Count Date	Eff	PFW	LCL	Mean	UCL
LB4110A - A1	Alpha	10/20/2015	10/22/2015	0.2314	P	0.2283	0.2320	0.2358
LB4110A - A2	Alpha	10/20/2015	10/22/2015	0.2112	P	0.2008	0.2127	0.2246
LB4110A - A3	Alpha	10/20/2015	10/22/2015	0.1998	P	0.1898	0.1969	0.2040
LB4110A - A4	Alpha	10/20/2015	10/22/2015	0.2257	P	0.2207	0.2247	0.2286
LB4110A - B1	Alpha	10/20/2015	10/22/2015	0.2154	P	0.2046	0.2115	0.2185
LB4110A - B2	Alpha	10/20/2015	10/22/2015	0.2150	P	0.2112	0.2191	0.2269
LB4110A - B3	Alpha	10/20/2015	10/22/2015	0.2400	P	0.2278	0.2365	0.2453
LB4110A - B4	Alpha	10/20/2015	10/22/2015	0.2325	P	0.2190	0.2272	0.2353
LB4110A - C1	Alpha	10/20/2015	10/22/2015	0.2069	P	0.2000	0.2070	0.2141
LB4110A - C2	Alpha	10/20/2015	10/22/2015	0.2223	P	0.1984	0.2184	0.2384
LB4110A - C3	Alpha	10/20/2015	10/22/2015	0.2506	P	0.2291	0.2492	0.2692
LB4110A - C4	Alpha	10/20/2015	10/22/2015	0.2205	P	0.2083	0.2170	0.2257
LB4110A - D1	Alpha	10/20/2015	10/22/2015	0.2151	P	0.2117	0.2202	0.2288
LB4110A - D2	Alpha	10/20/2015	10/22/2015	0.2421	P	0.2342	0.2410	0.2478
LB4110A - D3	Alpha	10/20/2015	10/22/2015	0.2510	P	0.2466	0.2571	0.2677
LB4110A - D4	Alpha	10/20/2015	10/22/2015	0.1948	P	0.1936	0.1945	0.1954
LB4110R - E1	Alpha	10/20/2015	10/22/2015	0.2342	P	0.2318	0.2337	0.2356
LB4110R - E2	Alpha	10/20/2015	10/22/2015	0.2090	P	0.2011	0.2120	0.2229
LB4110R - E3	Alpha	10/20/2015	10/22/2015	0.2124	P	0.2119	0.2126	0.2132
LB4110R - E4	Alpha	10/20/2015	10/22/2015	0.2447	P	0.2405	0.2462	0.2520
LB4110R - F1	Alpha	10/20/2015	10/22/2015	0.1910	P	0.1869	0.1926	0.1982
LB4110R - F2	Alpha	10/20/2015	10/22/2015	0.1988	P	0.1980	0.1991	0.2002
LB4110R - F3	Alpha	10/20/2015	10/22/2015	0.2407	P	0.2366	0.2399	0.2431
LB4110R - F4	Alpha	10/20/2015	10/22/2015	0.2219	P	0.2139	0.2202	0.2266
LB4110R - G1	Alpha	10/20/2015	10/22/2015	0.2052	P	0.1855	0.2009	0.2163
LB4110R - G2	Alpha	10/20/2015	10/22/2015	0.2043	P	0.2008	0.2056	0.2105
LB4110R - G3	Alpha	10/20/2015	10/22/2015	0.2296	P	0.2216	0.2279	0.2341
LB4110R - G4	Alpha	10/20/2015	10/22/2015	0.2029	P	0.2012	0.2026	0.2039

GPC Detector Report
(ALL Efficiencies)

Detector	Alpha/Beta	Calibration Date	Count Date	Eff	PFW	LCL	Mean	UCL
LB4110A - A1	Beta	10/20/2015	10/22/2015	0.5518	P	0.5401	0.5537	0.5674
LB4110A - A2	Beta	10/20/2015	10/22/2015	0.4747	P	0.4684	0.4821	0.4959
LB4110A - A3	Beta	10/20/2015	10/22/2015	0.4729	P	0.4643	0.4718	0.4793
LB4110A - A4	Beta	10/20/2015	10/22/2015	0.5335	P	0.5306	0.5398	0.5491
LB4110A - B1	Beta	10/20/2015	10/22/2015	0.5369	P	0.5225	0.5316	0.5408
LB4110A - B2	Beta	10/20/2015	10/22/2015	0.5240	P	0.5040	0.5235	0.5430
LB4110A - B3	Beta	10/20/2015	10/22/2015	0.5906	P	0.5596	0.5837	0.6078
LB4110A - B4	Beta	10/20/2015	10/22/2015	0.5564	P	0.5410	0.5548	0.5686
LB4110A - C1	Beta	10/20/2015	10/22/2015	0.5076	P	0.4569	0.4917	0.5264
LB4110A - C2	Beta	10/20/2015	10/22/2015	0.5255	P	0.4850	0.5250	0.5650
LB4110A - C3	Beta	10/20/2015	10/22/2015	0.6194	P	0.5750	0.6164	0.6578
LB4110A - C4	Beta	10/20/2015	10/22/2015	0.5501	P	0.5151	0.5358	0.5566
LB4110A - D1	Beta	10/20/2015	10/22/2015	0.6565	P	0.6518	0.6614	0.6709
LB4110A - D2	Beta	10/20/2015	10/22/2015	0.6313	P	0.6273	0.6375	0.6478
LB4110A - D3	Beta	10/20/2015	10/22/2015	0.6250	P	0.6143	0.6381	0.6620
LB4110A - D4	Beta	10/20/2015	10/22/2015	0.4764	P	0.4661	0.4763	0.4865
LB4110R - E1	Beta	10/20/2015	10/22/2015	0.5562	P	0.5546	0.5569	0.5592
LB4110R - E2	Beta	10/20/2015	10/22/2015	0.5162	P	0.5080	0.5145	0.5209
LB4110R - E3	Beta	10/20/2015	10/22/2015	0.5107	P	0.5106	0.5107	0.5108
LB4110R - E4	Beta	10/20/2015	10/22/2015	0.6073	P	0.5940	0.6124	0.6308
LB4110R - F1	Beta	10/20/2015	10/22/2015	0.4813	P	0.4673	0.4783	0.4893
LB4110R - F2	Beta	10/20/2015	10/22/2015	0.4799	P	0.4573	0.4751	0.4929
LB4110R - F3	Beta	10/20/2015	10/22/2015	0.6102	P	0.5590	0.5992	0.6394
LB4110R - F4	Beta	10/20/2015	10/22/2015	0.5488	P	0.5129	0.5411	0.5693
LB4110R - G1	Beta	10/20/2015	10/22/2015	0.4705	P	0.4371	0.4633	0.4894
LB4110R - G2	Beta	10/20/2015	10/22/2015	0.5066	P	0.5006	0.5053	0.5100
LB4110R - G3	Beta	10/20/2015	10/22/2015	0.5694	P	0.5060	0.5558	0.6055
LB4110R - G4	Beta	10/20/2015	10/22/2015	0.5101	P	0.4941	0.5066	0.5192